ALLEVYN◊ GENTLE BORDER LITE

SILICONE GEL ADHESIVE DRESSING | LOW ABSORBENCY | UP TO 7 DAYS WEAR TIME

Smith+Nephew

ALLEVYN◊ GENTLE BORDER LITE Silicone gel adhesive hydrocellular foam dressing

Description
ALLEVYN GENTLE BORDER LITE Dressings have a triple-layered construction which combines an absorbent hydrocellular pad sandwiched between a perforated silicone gel adhesive wound contact layer and a highly permeable waterproof top film.

Mode of action
ALLEVYN GENTLE BORDER LITE Dressings facilitate dynamic fluid management and promote a moist wound healing environment which is optimal for healing, thereby helping to reduce the risk of maceration. The perforated silicone gel adhesive wound contact layer allows exudate to pass through to the foam layer. The silicone adhesive layer is gentle, suitable for use on fragile skin and makes it easy to apply, reposition and remove the dressing. The hydrocellular foam layer absorbs exudate away from the wound. The breathable top film allows evaporation and aids the prevention of bacterial contamination of the wound.

Indications
Wound management by secondary intention on shallow, granulating wounds, chronic and acute exudative wounds, full and partial thickness wounds which include pressure ulcers, leg ulcers, diabetic foot ulcers, surgical wounds, first and second degree burns, skin tears. ALLEVYN GENTLE BORDER LITE Dressings can be used for the prevention of medical device related pressure injuries on intact skin, as part of a pressure injury protocol.

Precaution: Care should be taken when using the dressing under medical devices near the nose and mouth, where moisture can build up and affect adherence, leading to dressing movement.

Precautions
Do not use ALLEVYN GENTLE BORDER LITE Dressings with oxidising agents such as hypochlorite solutions (e.g. EUSOL) or hydrogen peroxide, as these can break down the absorbent polyurethane component of the dressing. If reddening or sensitisation occur discontinue use and consult a healthcare professional.
Do not use the dressings to secure other medical devices to the patient, due to the gentle adhesive used.
ALLEVYN GENTLE BORDER LITE Dressings should not be used for pressure injury prevention unrelated to medical devices.
Care should be taken when using the dressings under medical devices near the nose and mouth, where moisture can build up and affect adherence, leading to dressing movement.
ALLEVYN GENTLE BORDER LITE Dressings are single-use only and should not be re-used as cross-contamination or infection may occur and the dressing may be ineffective.
The scale on the pouch is for guidance to assist dressing selection only, and care should be taken to avoid cross-contamination.
ALLEVYN GENTLE BORDER LITE Dressings are for healthcare professional use and for use by patients under the direction of a healthcare professional.
When required, healthcare professionals should provide appropriate advice to patients on how to apply, monitor the dressing, remove, and dispose of the dressing and when to contact a healthcare professional as detailed in this leaflet.

Instructions for use
1. Cleanse the application site. Ensure surrounding skin is clean, dry and free from excess hair.
2. Select appropriate dressing size.
3. Start to remove the protector material from ALLEVYN GENTLE BORDER LITE Dressing and anchor the adhesive side of the dressing to the skin.
4. Smooth the dressing over the application site, removing the remaining protector material and ensure that the dressing is adhered all around the application site. ALLEVYN GENTLE BORDER LITE Dressings can be left in place without the need for secondary retention.
5. ALLEVYN GENTLE BORDER LITE Dressings can be cut, which help to dress awkward areas of the body. Cutting the dressing will compromise the bacterial barrier and showerproof properties of the dressing. Always use aseptic technique when cutting the dressing. Ensure any exposed foam areas are covered with an appropriate film dressing taking care not to cover the entire dressing.
6. ALLEVYN GENTLE BORDER LITE Dressings can be used in conjunction with INTRASITE◊ Gel for necrotic or sloughy wounds.

Observation when used for medical device related pressure injury prevention: frequently inspect the skin for pressure damage by lifting back one corner, inspecting, and then repositioning the dressing.

Frequency of change
During the early stages of treatment, ALLEVYN GENTLE BORDER LITE Dressings should be inspected frequently.
ALLEVYN GENTLE BORDER LITE Dressings can be left in place for up to 7 days depending on the condition of the application site and the surrounding skin or until exudate is visible and approaches to within 0.5cm/3/16in. of the edge of the dressing pad, whichever is sooner. The decision of when to change should be dependent upon clinical assessment and local protocols under the guidance of a healthcare professional

Dressing removal
To remove ALLEVYN GENTLE BORDER LITE Dressings, lift one corner of the dressing and slowly peel back until completely removed from the application site.
Dispose of according to local clinical protocols, or throw away hygienically after use.
If during the use of this device you believe that a serious incident has occurred, please report it to the manufacturer and your competent authority.
complaints@smith-nephew.com

Product Availability

Code	Size	Availability
66800833	5cm x 5cm / 2in. x 2in.	Carton of 10
66800834	7.5cm x 7.5cm / 3in. x 3in.	Carton of 10
66800835	10cm x 10cm / 4in. x 4in.	Carton of 10
66800836	5.5cm x 12cm / 2 1/8in. x 4 3/4in.	Carton of 10
66800837	8cm x 15cm / 3 1/8in. x 6in.	Carton of 10
66800840	15cm x 15cm / 6in. x 6in.	Carton of 10
66800841	8cm x 8cm / 3 1/8in. x 3 1/8in	Carton of 10
66801012	10cm x 20cm / 4in. x 8in.	Carton of 10
66801013	Oval 8.6cm x 7.7cm / 3 3/8in. x 3in.	Carton of 10
66802143	Oval 8.6cm x 7.7cm / 3 3/8in. x 3in.	Carton of 3
66801014	Oval 15.2cm x 13.1cm / 6in. x 5 1/8in.	Carton of 10
66801041	Multisite 8cm x 8.4cm / 3 1/8in. x 3 5/16in.	Carton of 10
66802144	Multisite 8cm x 8.4cm / 3 1/8in. x 3 5/16in.	Carton of 3

◊Trade Marks of Smith+Nephew. Date of issue 08/2024. Revision 02

ALLEVYN◊ GENTLE BORDER LITE



SILIKOONGEELIST KLEEPUV SIDE | MADAL IMAVUS | KANDMISAEG KUNI 7 PÄEVA

Smith+Nephew

ALLEVYN◊ GENTLE BORDER LITE Silikoongeelist adhesiivne hüdrotsellulaarne vahtplaaster

Kirjeldus
Sidemed ALLEVYN GENTLE BORDER LITE plaastrid on kolmekihilised, ühendades imava hüdrotsellulaarse padjakese, mis on paigutatud perforeeritud silikoongeelist haavavõrgu ja väga läbilaskva veekindla pealmise kile vahele.

Toimemehhanism
Sidemed ALLEVYN GENTLE BORDER LITE plaastrid hõlbustavad dünaamilist vedelikujaotust ja soodustavad haava paranemiseks optimaalset niisket keskkonda, aidates seeläbi vähendada matseratsiooniohtu. Perforeeritud silikoongeeli kleepuv haavavõrk võimaldab eksudaadil tungida läbi vahukihi. Kleepuv silikoonkiht on õrn, sobib kasutamiseks õrnal nahal ja võimaldab plaastrit hõlpsasti peale panna, ümber paigutada ja eemaldada. Hüdrotsellulaarne vahukiht imab eksudaadi haavast eemale. Hingav pealmine kile võimaldab aurustumist ja aitab vältida haava bakteriaalset saastumist.

Näidustused
Sekundaarne haavahooldus madalatele, granuleerivatele haavadele, kroonilistele ja ägedatele eksudeerivatele haavadele, täis- ja osalise paksusega haavadele, sh rõhuhaavad, jalahaavandid, diabeetiliste jalgade haavandid, kirurgilised haavad, esimese ja teise astme põletushaavad, naharebendid. Sidemeid ALLEVYN GENTLE BORDER LITE plaastreid võib kasutada tervel nahal meditsiiniseadmetega seotud survevigastuste ennetamiseks kooskõlas survevigastuste ravijuhendiga.

Ettevaatusabinõu: ettevaatlik tuleb olla, kui plaastrit kasutatakse meditsiiniseadmete all nina ja suu lähedal, kuhu võib koguneda niiskus ja mõjutada plaastri kinnitumist, mis võib põhjustada selle liikumist.

Ettevaatusabinõud
Ärge kasutage ALLEVYN GENTLE BORDER LITE'i plaastreid oksüdeerivate ainetega, näiteks hüpokloriidilahustega (nt EUSOL) või vesinikperoksiidiga, kuna need võivad lõhustada plaastri absorbeerivat polüuretaankomponenti. Kui tekib punetus või ülitundlikkus, lõpetage kasutamine ja konsulteerige tervishoiutöötajaga.
Ärge kasutage plaastreid muude meditsiiniseadmete kinnitamiseks patsiendi külge, kuna kasutatav liim on õrn.
Sidemeid ALLEVYN GENTLE BORDER LITE plaastreid ei tohi kasutada selliste survevigastuste ennetamiseks, mis ei ole seotud meditsiiniseadmetega.
Ettevaatlik tuleb olla, kui plaastrit kasutatakse meditsiiniseadmete all nina ja suu lähedal, kuhu võib koguneda niiskus ja mõjutada plaastri kinnitumist, mis võib põhjustada selle liikumist.
Sidemed ALLEVYN GENTLE BORDER LITE plaastrid on ühekordseks kasutamiseks ja neid ei tohi uuesti kasutada, kuna võib tekkida ristsaastumine või infektsioon ja plaaster võib muutuda ebaefektiivseks.
Pakendil olev skaala on abiks ainult plaastri valimisel ning ristsaastumise vältimiseks tuleb olla ettevaatlik.
ALLEVYN GENTLE BORDER LITE plaastrid on mõeldud kasutamiseks tervishoiutöötajatele ja patsientidele tervishoiutöötaja juhendamisel.
Vajaduse korral peaksid tervishoiutöötajad andma patsientidele asjakohast nõu, kuidas plaastrit peale panna, jälgida, eemaldada ja kõrvaldada ning millal võtta ühendust tervishoiutöötajaga, nagu on kirjeldatud käesolevas infolehes.

Kasutusjuhised
1. Puhastage paigalduskoht. Veenduge, et ümbritsev nahk oleks puhas, kuiv ja ilma liigsete karvadeta.
2. Valige sobiv haavasideme suurus.
3. Alustage plaastri ALLEVYN GENTLE BORDER LITE kaitsematerjali eemaldamist ja kinnitage plaastri kleepuv pool naha külge.
4. Tasandage plaaster üle paigalduskoha, eemaldades ülejäänud kaitsematerjali ja tagades, et plaaster kinnituks kogu paigalduskoha ümber. Sidemeid ALLEVYN GENTLE BORDER LITE plaastri võib jätta peale ilma sekundaarse kinnituseta.
5. Sidemeid ALLEVYN GENTLE BORDER LITE plaastreid saab lõigata – need aitavad katta kehapiirkondi, mida on muidu raske katta. Plaastri lõikamine kahjustab plaastri bakteriaalset barjääri ja veekindlaid omadusi. Kasutage plaastri lõikamisel alati aseptilist tehnikat. Veenduge, et kõik avatud vahtkattega alad on kaetud sobiva kilekattega, jälgides, et see ei kataks kogu plaastrit.
6. Sidemeid ALLEVYN GENTLE BORDER LITE plaastrit võib kasutada koos geeliga INTRASITE GEL nekrootiliste või lahtiste haavade puhul.

Vaatlus, kui seda kasutatakse meditsiiniseadmetega seotud survevigastuste vältimiseks: kontrollige sageli naha survevigastusi, tõstes ühe nurga üles, kontrollides ja seejärel asetades plaastri tagasi oma kohale.

Vahetamise tihedus
Ravi algstaadiumis tuleb ALLEVYN GENTLE BORDER LITE plaastreid sageli kontrollida.
Sidemeid ALLEVYN GENTLE BORDER LITE plaastri võib jätta peale kuni 7 päevaks olenevalt paigalduskoha ja ümbritseva naha seisundist või kuni eksudaat on nähtav ja läheneb kuni 0,5 cm kaugusele plaastripadja servast, olenevalt sellest, kumb toimub varem. Otsus, millal plaaster vahetada, tuleks teha tervishoiutöötaja juhendamisel ja see peaks põhinema kliinilisel hinnangul ja kohalikel ravijuhenditel.

Plaastri eemaldamine
Sidemete ALLEVYN GENTLE BORDER LITE plaastrite eemaldamiseks tõstke selle üks nurk üles ja tõmmake plaastrit aeglaselt tagasi, kuni see on täielikult eemaldatud.
Kõrvaldage vastavalt kohalikele kliinilistele ravijuhenditele või visake pärast kasutamist hügieeniliselt minema.
Kui te usute, et selle seadme kasutamise ajal on esinenud ohujuhtum, teatage palun sellest tootjale ja oma pädevale riigiasutusele.
complaints@smith-nephew.com

Toote saadavus

Kood	Suurus	Saadavus
66800833	5 cm x 5 cm	10 tk karbis
66800834	7,5 cm x 7,5 cm	10 tk karbis
66800835	10 cm x 10 cm	10 tk karbis
66800836	5,5 cm x 12 cm	10 tk karbis
66800837	8 cm x 15 cm	10 tk karbis
66800840	15 cm x 15 cm	10 tk karbis
66800841	8 cm x 8 cm	10 tk karbis
66801012	10 cm x 20 cm	10 tk karbis
66801013	Ovaalne side 8,6 cm x 7,7 cm	10 tk karbis
66802143	Ovaalne side 8,6 cm x 7,7 cm	3 tk karbis
66801014	Ovaalne side 15,2 cm x 13,1 cm	10 tk karbis
66801041	Mitme asukoha side 8 cm x 8,4 cm	10 tk karbis
66802144	Mitme asukoha side 8 cm x 8,4 cm	3 tk karbis

◊Smith+Nephew kaubamärgid.
Väljaandmise kuupäev 08/2024. Läbivaatamine 02.